


PULSE GENERATOR

PERFORMANCE CHECK

Model: *APR-B3-0A-P*

S.N.: *6027*

Date: *SEPT 10 1991*



- a) Output signal amplitude:
0 TO +100 VOLTS
- b) Pulse width: *(TO 50ns)*
20 TO 200 NS
- c) Rise time:
≤ 3 NS
- d) Fall time:
≤ 3 NS
- e) PRF: *0 TO 20 KHz*
- f) Jitter, stability:
OK
- g) Prime power:

50 db ATTEN
∴ ≈ 32 VOLTS/DIV
PRF ≈ 20 KHz
R_L = 50 Ω.

+24 VOLTS
300 mA (MAX)