


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVL-2-PS-P-M

S.N.: 6055

Date: OCT 2 1991



70 dB ATTEN
320 VOLTS/DIV
PRF = 2 KHZ

- a) Output signal amplitude:
0 TO +350 VOLTS
- b) Pulse width: 70 50 ns
- c) Rise time: 5 TO 100 NS
≤ 2 NS
- d) Fall time:
≤ 3 NS
- e) PRF: 0 TO 5 KHZ
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 HZ.

[Signature]