


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVP-M-1-M-PN*

S.N.: *6062*

Date: *OCT 18 1991*



*40 dB ATTEN*

*10 VOLTS/DIV*

*1MHz 700kHz*

- a) Output signal amplitude:  
*0 TO ± 10 VOLTS*
- b) Pulse width:  
*0.2 TO 40 NS*
- c) Rise time:  
*≤ 100 PS*
- d) Fall time:  
*≤ 130 PS*
- e) PRF:  
*0 TO 100 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*15 VOLTS*  
*250 mA (MAX)*

*[Signature]*