


PULSE GENERATOR

PERFORMANCE CHECK

Model: *MR-4B-C-P-N14A*


S.N.: *6090*

Date: *NOV 21 1991*



- a) Output signal amplitude:
0 TO +400 VOLTS
- b) Pulse width: *TO $R_L \geq 10K$
 $\approx 100NS$ TO $10 \mu S$*
- c) Rise time:
65NS
- d) Fall time:
 $\leq 30 NS$
- e) PRF: *50 Hz TO 5 KHz*
- f) Jitter, stability:
OK
- g) Prime power:
*120/240 V
50-60 Hz*

Ⓐ $R_L \geq 10 K$
50 NS / DIV
PRF = 5.0 KHz } *NARROW PULSE*



Ⓑ $R_L \geq 10K$
1.0 μS / DIV
PRF = 5.0 KHz } *WIDE PULSE*