


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVL-2 PS PN-T


S.N.: 6104

Date: NOV 20 1991



- a) Output signal amplitude:
0 TO ± 350 VOLTS
- b) Pulse width: TO 50ns
- c) Rise time: 5 TO 100 NS
2 TO 10 NS
- d) Fall time: 3 TO 10 NS
- e) PRF: 0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
5.0-60 Hz

(A) P_{out} TR MIN
70 db ATTEN
 ≈ 320 VOLTS/DIV
PRF = 3 KHz



(B) N_{out} TR MAX.
70 db ATTEN
 ≈ 320 VOLTS/DIV
PRF ≈ 3 KHz

[Handwritten signature]