


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVD-D-4-C-41A-R4-PNS*

S.N.: *6106*

Date: *DEC 10 1991*



A)

B)

- a) Output signal amplitude: *0 TO 20 VOLTS TO*
- b) Pulse width: *100 µs*
- c) Rise time: *10 TO 100 µs*
- d) Fall time: *≤ 5 µs*
- e) PRF: *0 TO 5 KHz*
- f) Jitter, stability: *OK*

A) 5 VOLTS / DIV

B) 20 VOLTS / DIV

R_L = 100 Ω

- g) Prime power: *120/240 V*
50-60 Hz
- h) *M: 30 TO 300 ns*
- i) *AS: 5 TO 500 ns*
- j) *OS: 0 TO +5 VOLTS*