


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVD-2-AW-P-IGIA*

S.N.: *6118*

Date: *JAN 20 1992*



*≈ 6.5 Amps / DIV*  
*PRF = 10 KHz.*

- a) Output signal amplitude:  
*0 TO 15 Amps*
- b) Pulse width:  
*2 TO 100 NS*
- c) Rise time:  
*≤ 1 NS*
- d) Fall time:  
*≤ 2 NS*
- e) PRF:  
*0 TO 10 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*+15 Volts.*  
*300 mA (max)*  
