


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *MR-3A-PW-C-P*

S.N.: *6127*

Date: *NOV 28 1991*



- a) Output signal amplitude: *0 TO +200 V TO*
- b) Pulse width: *50 ns*
- c) Rise time: *0.1 TO 5.0 ns*  
*(0.5% MAX DUTY CYCLE)*
- d) Fall time: *≤ 10 ns*
- e) PRF: *10 Hz TO 10 KHz*  
*(0.5% MAX DUTY)*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240 V*  
*50-60 Hz*

*R<sub>L</sub> = 50 Ω*  
*PRF = 1 kHz*