


PULSE GENERATOR

PERFORMANCE CHECK


Model: *MT-V-1403-C-P*

S.N.: *6138*

Date: *JAN 24 1992*



Ⓐ 60 db ATTEN
∴ 100 VOLTS/DIV
RATE = 2 KHZ



Ⓑ 60 db ATTEN
1 NS/DIV
(RISE TIME)

- a) Output signal amplitude:
0 TO +100 VOLTS
- b) Pulse width:
5 TO 100 NS
- c) Rise time:
≤ 200 PS
- d) Fall time:
≤ 3 NS
- e) PRF:
0 TO 5 KHZ
- f) Jitter, stability:
OK
- g) Prime power:
120/240 VOLTS
50-60 HZ.