


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *M6-1-P*  
S.N.: *6140*  
Date: *JAN 22 1992*



- a) Output signal amplitude:  
*0 TO +250 VOLTS*
- b) Pulse width:  
*≤ 4 NS*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 TO 50 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*+15 VOLTS*  
*250 mA (MAX)*

*60 db ATTEN*  
*100 VOLTS/DIV*  
*PRF = 30 KHz*

*[Signature]*