


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *PH-100PT-F*

S.N.: *6150*

Date: *JAN 30 1992*



*40 db ATTEN*  
*10 VOLT/DIV*  
*700 KHZ*

- a) Output signal amplitude:  
*≥ 30 VOLTS*
- b) Pulse width:  
*≤ 1 NS*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 TO 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*+15 VOLTS*  
*100 mA (MAX)*  
*[Signature]*