



PULSE GENERATOR
PERFORMANCE CHECK


Model: MG-2-C-P-SNLI-TPI-RF
 S.N.: 6179
 Date: FEB 19 1992



RANGE 1 60 db ATTEN
 - 100 VOLTS/DIV
 PRF = 10 kHz




RANGE 2 60 db ATTEN
 - 100 VOLTS/DIV
 PRF = 10 kHz



RANGE 3 60 db ATTEN
 - 100 VOLTS/DIV
 PRF = 10 kHz

- a) Output signal amplitude:
0 TO +100 VOLTS
- b) Pulse width: TO 50ns
1) 2 NS
2) 4 NS
- c) Rise time: 3) 10 NS
4) 20 NS
- d) Fall time:
- e) PRF: 10 Hz TO 10 kHz
- f) Jitter, stability:
OK
- g) Prime power:

120 / 240 V
 50 - 60 Hz



RANGE 4 60 db ATTEN
 - 100 VOLTS/DIV
 PRF = 10 kHz

[Handwritten signature]