


PULSE GENERATOR

PERFORMANCE CHECK


Model: *AV6-2-C-P-SNL1-TPI-R4*

S.N.: *6182*


Date: *FEB 26 19 92*



RANGE 1 100V/DIV



RANGE 2 100V/DIV



RANGE 3 100V/DIV

a) Output signal amplitude:  
*0 TO +100 VOLTS*

b) Pulse width:  
*2, 4, 10, 20 NS*

c) Rise time:  
*—*


d) Fall time:  
*—*

d1)  $t_p \leq 40 NS$

e) PRF: *10 Hz TO 10 KHz*

f) Jitter, stability:  
*OK*

g) Prime power:  
*120/240 V*  
*50-60 Hz*



RANGE 4 100V/DIV