


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVH-14V1-PN-55C1


S.N.: 6185

Date: FEB 6 1992



- a) Output signal amplitude:  
± 100 VOLTS TO 50V
- b) Pulse width:  
P: 0.6 TO 2.0 NS  
N: 0.6 TO 2.0 NS
- c) Rise time:  
NA
- d) Fall time:  
NA
- e) PRF: 0 TO 100 K/HZ
- f) Jitter, stability:  
OK
- g) Prime power:  
+15 VOLTS DC  
150 mA (MAX)

Point 60db ∴ 100 VOLTS/DIV



Point 60db ∴ 100 VOLTS/DIV