


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVO-5A-e-P*

S.N.: *6189*

Date: *FEB 11 1992*



- a) Output signal amplitude:
0 TO 40 AMP
- b) Pulse width: (*0.75 div*)
20 TO 150 NS
- c) Rise time:
≤ 5 NS
- d) Fall time:
≤ 5 NS
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:

$$R_c = 0.7 \Omega$$
$$\approx 14 \text{ AMP/DIV}$$

$$\text{PRF} = 1.0 \text{ KHz}$$

$$120 / 240 \text{ V}$$
$$50 \sim 60 \text{ Hz}$$

[Signature]