


PULSE GENERATOR

PERFORMANCE CHECK


Model: *APR-A-1-PW-C-05-PN-PWT-AT*

S.N.: *6193*

Date: *FEB 24 1992*



*P_{out}: PRF = 50 KHz
R_L = 50 Ω*



*N_{out}: PRF = 5 KHz
R_L = 50 Ω*

- a) Output signal amplitude: *0 TO ± 200 VOLTS*
- b) Pulse width: *TO 50 ns*
- c) Rise time: *50 NS TO 1.0 μ S
(0.5% MAX DUTY CYCLE)*
- d) Fall time: *≤ 10 NS*
- e) PRF: *≤ 10 NS
0 TO 100 KHz
(0.5% MAX DUTY)*
- f) Jitter, stability: *CYCLE*
- g) Prime power: *OK
120/240 VOLTS
3.0-60 Hz*

[Signature]