


PULSE GENERATOR

PERFORMANCE CHECK


Model: AVP-2S-P-0S-ECL

S.N.: 6199

Date: FEB 18/92



A) 20 db ATTEN
∴ 2 VOLTS/DIV
PRF = 1.0 MHz



B) AS A) BUT 50 PS/DIV
(RISE TIME)

$$T_{RT} = \sqrt{T_{RAS}^2 + T_{SCOPE}^2 + T_{ATTEN}^2}$$

$$\approx 50 \text{ PS (20-80\%)}$$

$$T_{SCOPE} = 25 \text{ PS}$$

18 GHz ATTEN + CABLES

- a) Output signal amplitude:
0 TO +2 VOLTS
- b) Pulse width:
0.2 TO 4 NS
- c) Rise time:
 $\leq 40 \text{ PS}$
- d) Fall time:
 $\leq 135 \text{ PS}$
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 VOLTS
400 mA (MAX)

∴ CONCL THAT

$$T_{RAS} \leq 40 \text{ PS}$$