


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVL-AV-1-C-W-P

S.N.: 6205

Date: FEB 20 1992



60 dB ATTEN

100 VOLTS/DIV

PRF = 5.0 KHz

- a) Output signal amplitude:  
0 TO 4100 VOLTS
- b) Pulse width:  
3 TO 400 NS\*
- c) Rise time:  
≤ 1 NS
- d) Fall time:  
≤ 2 NS
- e) PRF: 0 TO 5 KHz
- f) Jitter, stability:  
OK
- g) Prime power:  
120/240 V  
50 = 60 Hz

*[Signature]*