



PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVI-V-C-P*
S.N.: *6235*
Date: *APRIL 10 1992*



Ⓐ 50 dB ATTEN
32 VOLTS/DIV
RIF ≈ 20 KHz
(RISE TIME)



Ⓑ 10ns Ⓐ BAT
10 NS/DIV

- a) Output signal amplitude:
0 TO +50 VOLTS
- b) Pulse width: *TO 50ns*
2 TO 100 NS
- c) Rise time:
≤ 350 PS
- d) Fall time:
≤ 2 NS
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
O.K.
- g) Prime power:
120/240V