


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-6A-C-P-CA-CW-LIB*

S.N.: *6238*

Date: *APRIL 16 1992*



- a) Output signal amplitude:
0 TO +1.0 AMP
- b) Pulse width:
50 NS TO 5.0 μ S
- c) Rise time:
 ≤ 10 NS
- d) Fall time:
 ≤ 10 NS
- e) PRF:
0 TO 1.0 KHz
- f) Jitter, stability:
OK

- g) Prime power:
120/240 V
50-60 Hz

*MANUFACTURER
OUT TO $R_L = 50 \Omega$*

[Signature]