


PULSE GENERATOR

PERFORMANCE CHECK


Model: AVR-A-1-S1-P5-PN

S.N.: 6274

Date: MAY 27 1992



P_{out} : 60 dB ATTEN
- 100 VOLT/DIV
PRF = 10 KHz



N_{out} : $R_L = 50 \Omega$
PRF = 5 KHz

- a) Output signal amplitude:
0 TO ± 200 VOLTS
- b) Pulse width: 40 50 ns
50 NS TO 10 μ S
- c) Rise time: (0.5% MAX DUTY)
CYCLE
 ≤ 2 NS
- d) Fall time:
 ≤ 2 NS
- e) PRF: 0 TO 100 KHz
(0.5% MAX DUTY)
- f) Jitter, stability: CYCLE
OK
- g) Prime power:
120/240 V
50-60 Hz