


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AVH-HV.1-PN-SSCJ*

S.N.: *6279*

Date: *APRIL 28 92*



*F_{out} 60 dB
- 100VOLT*



N_{out} AS P.

- a) Output signal amplitude:
± 100 VOLT?
- b) Pulse width:
0.6 TO 2.0NS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 100 K/HZ
- f) Jitter, stability:
OK
- g) Prime power:
*+15 VOLTS DC
150mA MAX.*

h) SPURIOUS.

*2 x 10 LOWER
THAN FIRST
8 UNITS IN
THIS SERIES*

MJ