


PULSE GENERATOR
PERFORMANCE CHECK


Model: *APP-2-C-P-1M*

S.N.: *6286*

Date: *JUNE 5 1992*



*A) RANGE
40 db ATTEN
10 VOLTS/DIV*



*B) RANGE
40 db ATTEN
10 VOLTS/DIV
PRF = 100 KHz*

- a) Output signal amplitude:
0 TO + 20 VOLTS
- b) Pulse width: *TO 500*
A) 0.4 TO 8.0 NS
B) 8 TO 100 NS
- c) Rise time:
≤ 200 PS
- d) Fall time:
≤ 300 PS
- e) PRF:
0 TO 100 KHz
- f) Jitter, stability:
OK
- g) Prime power:

*120/200 V
50-60 KHz*
[Signature]