


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVI-F*  
S.N.: *6330*  
Date: *JUNE 18 1992*



*50 dB ATTEN*  
*~32 VOLTS/DIV*  
*PRF = 50 KHz*

- a) Output signal amplitude:  
*0 TO +40 VOLTS*
- b) Pulse width: *TO 50ns*
- c) Rise time: *2 TO 100ns*  
*≤ 350 ps*
- d) Fall time: *≤ 2 ns*
- e) PRF: *MIN PW: 0 TO 1.0ms*
- f) Jitter, stability: *MAX PW: 5 TO 20 KHz.*  
*OK*
- g) Prime power: *+15 VOLTS.*  
*150mA (max)*