


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AVMP-4-C-P-R4-TIA-03*

S.N.: *6331*


Date: *JUNE 26 1992*



*A CHANNEL, P_{out}.
40dB ATTEN
10 VOLTS/DIV*



*A CHANNEL N_{out}.
40 dB
10 VOLTS/DIV*



- a) Output signal amplitude:
0 TO ± 10 VOLTS
- b) Pulse width: *TO 50 ns
10 NS TO 10 μS
(10% MAX DUTY CYCLE)*
- c) Rise time:
≤ 200 ps
- d) Fall time:
≤ 200 ps
- e) PRF:
*0 TO 1.0 MHz
(10% MAX DUTY CYCLE)*
- f) Jitter, stability:
OK
- g) Prime power:
*120/240 VOLTS
30 & 60 Hz.*

*1 B CHANNEL
P_{out}.
R_L = 50Ω
10% DUTY CYCLE*