


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVH-MPT-TPI-P*

S.N.: *6336*

Date: *MAY 30 1992*



40 db ATTEN

10 VOLTS/DIV

PRF = 1.0 MHz

- a) Output signal amplitude:
≈ + 30 VOLTS
- b) Pulse width:
≤ 1 ns
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
*+15 VOLTS,
100 mA (MAX)*
[Signature]
- h) t_p : *≤ 2 ns.*