


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVR-A-1-TR-C-7*


S.N.: *6341*

Date: *JUNE 8 1992*



- a) Output signal amplitude:
0 TO + 200 VOLTS
- b) Pulse width: *TO 50ns*
50ns TO 1.0us
- c) Rise time: *(0.5% MAX DUTY CYCLE)*
- d) Fall time: *10 TO 40 ns*
- e) PRF: *0 TO 100 KHz*
(0.5% MAX DUTY)
- f) Jitter, stability: *(CYCLE)*
OK
- g) Prime power:
120 / 240 Vrms
50-60 Hz.

Ⓐ MIN RISE TIME
 $R_L = 50\Omega$
PRF = 10 KHz



Ⓑ MAX RISE TIME
 $R_L = 50\Omega$
PRF = 10 KHz

[Signature]