

PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVMP-2-C-P*

S.N.: *6352*

Date: *JUNE 15 1992*

a) Output signal amplitude:

0 TO +10 VOLTS

b) Pulse width: *TO 50 ns*

5 TO 100 NS

c) Rise time:

≤ 100 PS

d) Fall time:

≤ 135 PS

e) PRF:

0 TO 10 MHz


f) Jitter, stability:

OK

g) Prime power:

120/240 VOLTS

50-60 MHz



RISE TIME
1 NS/DIV

40 db ATTEN

10 VOLTS/DIV

PRF = 10 MHz

[Signature]