


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVR-5A-PS-AT-PWT-N

S.N.: 6353

Date: JUL 21 1992



$R_L = 50 \Omega$
100 VOLTS/DIV
PRF = 7.0 KHz

- a) Output signal amplitude:
0 TO -500 VOLTS
- b) Pulse width: ($R_L \geq 50 \Omega$)
0.1 TO 5.0 μ S
- c) Rise time: ($\leq 1\%$ MAX DUTY CYCLE)
 ≤ 30 NS
- d) Fall time:
 ≤ 30 NS
- e) PRF: 0 TO 10 KHz
($\leq 1\%$ MAX DUTY CYCLE)
- f) Jitter, stability:
OK
- g) Prime power:

120 / 240 V
50 / 60 Hz