

PULSE GENERATOR

PERFORMANCE CHECK

Model:


AVNC-1-C-TA

S.N.:

6393

Date:

JUL 29 1992



a) Output signal amplitude:

+ 5 Volts, $R_L \geq 50 \Omega$

b) Pulse width:

DUTY CYCLE FIXED

AT 50%

$\approx 50 \text{ MHz}$)

Rise time:

$\leq 5 \text{ NS}$

d) Fall time:

$\leq 5 \text{ NS}$

$\approx 6 \text{ MHz}$

(50 ns/div)

PRF:

1 kHz TO

50 MHz

f) Jitter, stability:

OK

g) Prime power:

120 / 240 Volts,
50-60 Hz

$P_{in} = 50 \text{ W}$