


PULSE GENERATOR
PERFORMANCE CHECK

Model: *MV0-6A-C-P-EP-EW-LIB*


S.N.: *6404*

Date: *SEPT 9 1992*



- a) Output signal amplitude:
0 TO +100 AMP
- b) Pulse width:
50 NS TO 5.0 US
- c) Rise time:
≤ 10 NS
- d) Fall time:
≤ 10 NS
- e) PRF:
0 TO 1.0 KHz.
- f) Jitter, stability:
OK

Ⓐ MAINFRAME OUT
TO $R_L = 50 \Omega$
(200 MHz SCOPE)



- g) Prime power:
120/240 V
50-60 Hz.
ref

