


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVL-3A-P-LCLA*

S.N.: *6412*

Date: *APR 6 1992*



*70 db ATTEN*

*≈ 320 VOLTS/DIV*

*PRF = 10 KHz*

- a) Output signal amplitude: *≈ ±300 VOLTS. DIV*
- b) Pulse width: *50n*
- c) Rise time: *10 nS*
- d) Fall time: *≈ 600 pS*
- e) PRF: *0 TO 20 KHz*
- f) Jitter, stability: *OK*
- g) Prime power: *±15 ±1.0 VOLTS*  
*250 mA (MAX)*