


PULSE GENERATOR

PERFORMANCE CHECK

Model: *M6-3A-COB1*

S.N.: *6446*

Date: *OCT 19 1992*



*↑ A      ↑ B*

*60 dB ATTEN*

*100 VOLTS/DIV*

*PRF = 20 KHz.*

- a) Output signal amplitude:  
*+200 TO +450 VOLTS*
- b) Pulse width:  
*≈ 8 NS*
- c) Rise time:  
*≈ 3 NS*
- d) Fall time:  
*≈ 3 NS*
- e) PRF: *0 TO 20 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*+15 VOLTS  
200 mA MAX.*