


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVG-3-P*
S.N.: *6450*
Date: *Aug 31 1992*



- a) Output signal amplitude:
+150 TO 500 VOLTS
- b) Pulse width: *(TO 50ns)*
≤ 5 NS (0.5 power)
- c) Rise time:
≤ 2 NS
- d) Fall time:
≤ 10 NS
- e) PRF: *0 TO 20 KHz*
- f) Jitter, stability:
OK
- g) Prime power:

60 db ATTEN
100 VOLTS/DIV
PRF = 10 KHz.

+15 VOLTS,
300 mA (MAX)