


PULSE GENERATOR

PERFORMANCE CHECK


Model: *MH-5-1-P-13LA*

S.N.: *6462*

Date: *NOV 12 1992*



ⓐ 40 dB ATTEN
∴ 10 VOLTS/DIV
PRF = 300 KHz



ⓑ AS @ OUT
100 PS/DIV

- a) Output signal amplitude: *> 17 VOLTS TO 50Ω*
- b) Pulse width: *(≈ 11 VOLTS)
≈ 80 NS FWHM*
- c) Rise time: *(COULD NOT MEASURE
RISE TIME)*
- d) Fall time: *NA*
- e) PRF: *0 TO 1.0 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *+15 VOLTS
150 mA (MAX)*