


PULSE GENERATOR
PERFORMANCE CHECK

Model: AV-145
S.N.: 6466
Date: OCT 12 1992



- a) Output signal amplitude:
0 TO ± 10 VOLTS TO
- b) Pulse width: $R_L > 50 \Omega$
5 NS \rightarrow DC
- c) Rise time:
 ≤ 3 NS
- d) Fall time:
 ≤ 3 NS
- e) PRF:
DC - 100 MHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 VOLTS
100 mA

INPUT: +5 VOLTS
LOAD: 50 OHMS
SCOPE: 200 MHz BW,
5 VOLTS/DIV
PRF: 1 MHz