


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV14-S-1-P*
S.N.: *6479*
Date: *OCT 4 1992*



- a) Output signal amplitude:
0 TO +10 VOLTS
- b) Pulse width:
≤ 130 PS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

40 db ATTEN
10 VOLTS/DIV
PRF ≈ 500 kHz

+15 VOLTS
125 mA.

[Signature]