


PULSE GENERATOR

PERFORMANCE CHECK


Model: AVR-61-C-P-IBM1*

S.N.: 6499

Date: NOV 11 1992



(A)



(B)

↑
RISE TIME.

- a) Output signal amplitude:
0 TO +100 VOLTS TO
- b) Pulse width: $R_L > 10K$
100 NS TO 1.0 MS
- c) Rise time: $\leq 10 NS$
- d) Fall time: $\leq 10 NS$
- e) PRF: 10 Hz TO 100 KHz
(50% MAX DUTY CYCLE)
- f) Jitter, stability:
OK
- g) Prime power:
120/240 VOLTS
50-60 Hz

* IBM1 OPTION:

$V_{out} \rightarrow 100 VOLTS$
 $PRF_{max} \rightarrow 100 KHz$
 $DUTY CYCLE \rightarrow 50\%$
MAX