


PULSE GENERATOR

PERFORMANCE CHECK

Model: AV6-3A-CO A1

S.N.: 6520

Date: DEC 10 92



- a) Output signal amplitude:  
+150 TO +450 VOLTS
- b) Pulse width:  
7 ns (AT 20% POINT)
- c) Rise time:  
≈ 3 ns
- d) Fall time:  
≈ 3 ns
- e) PRF:  
0 TO 20 KHz
- f) Jitter, stability:  
OK
- g) Prime power:  
+15 VOLTS  
350 mA (RM MS)
- h) SPURIOUS  
AS PER  
COBI UNITS.

60 db ATTEN  
100 VOLTS/DIV  
PRF ≈ 10 KHz