


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVG-3A-CO1*

S.N.: *6522*

Date: *DEC 10/92*



*60 db ATTEN
100 VOLTS/DIV.
PRF ≈ 10 KHz*

- a) Output signal amplitude:
+ 150 TO +450 VOLTS.
- b) Pulse width:
7 NS (AT 20% POINT)
- c) Rise time:
≈ 3 NS
- d) Fall time:
≈ 3 NS
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:
*+ 15 VOLTS
350 mA (MAX)*
- h) Spurious:
As per COB1 units.