


PULSE GENERATOR

PERFORMANCE CHECK

Model: AV6-3A-C0B1

S.N.: 6535

Date: JAN 22 1993



- a) Output signal amplitude:  
+150 TO +450 VOLTS
  - b) Pulse width:  
≈ 8 NS
  - c) Rise time:  
≈ 3 NS
  - d) Fall time:  
≈ 3 NS
  - e) PRF:  
0 TO 20 KHz
  - f) Jitter, stability:  
OK
  - g) Prime power:  
+15 VOLTS  
1.0 AMP (MAX)
- [Signature]*

↑ A      ↑ B  
60 dB ATTEN  
- 100 VOLTS/DIV  
PRF = 10 KHz