


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PK-A-P*

S.N.: *6558*

Date: *FEB 10 1993*



- a) Output signal amplitude:
0 TO +15 V
- b) Pulse width:
4 TO 100 NS
- c) Rise time:
≤ 100 PS
- d) Fall time:
62 NS
- e) PRF:
MIN PW: 0 TO 1.0 MHz
MAX PW: 0 TO 20 KHz
- f) Jitter, stability:
OK

40 db ATTEN

10 VERTICAL/DIV

PRF = 100 KHz

- g) Prime power:
+15 VOLTS
200 mA (max)

[Signature]