


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVI-P-PS-KOMP1

S.N.: 6570

Date: FEB 12 1993



- a) Output signal amplitude:
0 TO +20 VOLTS TO
- b) Pulse width: 50 nS
2 TO 50 NS
- c) Rise time:
≤ 400 PS
- d) Fall time:
≤ 2 NS
- e) PRF: 0 TO 20 KHz
- f) Jitter, stability:
OK

40 dB ATTEN

10 VOLTS/DIV

PRF = 20 KHz

- g) Prime power:
120/240 V
50-60 MHz

h) PROPAGATION DELAY:
≤ 5 NS.