


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVO-2A-P-5mA/AX-S3A*

S.N.: *6574*

Date: *FEB 24 1993*



Ⓐ  
← AX-S3A  
WPI OUT

Ⓑ  
← AVO-2A  
OUT  
(100mV)

- a) Output signal amplitude:  
*0 TO +5 AMPS*
- b) Pulse width:  
*2 TO 50 NS*
- c) Rise time:  
*≤ 2 NS*
- d) Fall time:  
*≤ 1 NS*
- e) PRF:  
*0 TO 20 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

Ⓐ 40 dB ATTEN  
∴ 5 VOLTS/DIV  
∴ 5 AMPS

+15 VOLTS  
250 mA (MMA)

Ⓑ 60 dB ATTEN  
∴ 100 VOLTS/DIV

*[Signature]*