


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV6-1-P-LT2B*

S.N.: *6599*

Date: *MAY 18 1993*



60 db ATTEN
100 VOLTS/DIV
PRF = 10 KHZ.

- a) Output signal amplitude: *± 250 VOLTS TO*
- b) Pulse width: *50 ns*
≤ 4 ns
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 TO 50 KHZ.*
- f) Jitter, stability: *OK*
- g) Prime power: *+15 VDC*
150 mA (MAX)

[Signature]