


PULSE GENERATOR

PERFORMANCE CHECK

Model: *MP-M-1403-C-P*

S.N.: *6648*

Date: *APRIL 30 1993*



- a) Output signal amplitude:
0 TO 40 VOLTS
- b) Pulse width:
0.4 TO 2.0 NS
- c) Rise time:
≤ 150 PS
- d) Fall time:
≤ 250 PS
- e) PRF: *0 TO 1.0 MHz*
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz

50 dB ATTEN
32 VOLTS/DIV
PRF = 100 kHz