


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVO-7C-CP-PN-CA-PWT*

S.N.: *6674*


Date: *JUNE 8 '93*



- a) Output signal amplitude:  
*0 to ±30 Amps*  
*0 to ±30 Volts*
- b) Pulse width:  
*0.2 to 50 μs*
- c) Rise time: *(0.25% max duty cycle)*  
*≤ 50 ns*
- d) Fall time:  
*≤ 50 ns*
- e) PRF:  
*0 to 10 kHz*  
*(0.25% max duty cycle)*
- f) Jitter, stability: *(cycle)*

*Post*  $R_L = 1 \Omega$  ( $\therefore 10 \text{ Amp/div}$ )  
*PRF = 5 kHz*

- g) Prime power:  
*120/240 V*  
*50-60 Hz.*



*Next*  $R_L = 1 \Omega$  ( $\therefore 10 \text{ amp/div}$ )  
*PRF = 50 kHz.*

*[Signature]*