


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AMP-AN-4V3-N*

S.N.: *6675*

Date: *JUNE 30 93*



50 Ohm ATTEN

32 VOLT/DIV

PRF = 100 kHz

- a) Output signal amplitude:
0 to -40 V to
- b) Pulse width: *50 n*
0.4 to 2.0 nS
- c) Rise time:
≤ 150 pS
- d) Fall time:
≤ 250 pS
- e) PRF:
0 to 100 kHz
- f) Jitter, stability:
OK
- g) Prime power:
175 VOLT
250 mA

[Signature]