


PULSE GENERATOR

PERFORMANCE CHECK

Model: *ANMH-4-P-BNWA*

S.N.: *6689*

Date: *JULY 14 1993*



a) Output signal amplitude:

0 TO 100 VOLTS

b) Pulse width: *(TO 50ns)*

3-4 NS

c) Rise time:

≈ 1.5 NS

d) Fall time:

≈ 1.5 NS

e) PRF:

0 TO 100 MHz

f) Jitter, stability:

OK

g) Prime power:

+24 VOLTS

*400 mA
(max)*

60 dB ATTEN

100 VOLTS/DIV