


PULSE GENERATOR

PERFORMANCE CHECK

Model: AV-155-PS-LPRA

S.N.: 6697

Date: JUNE 22 1993



- a) Output signal amplitude: 0 TO +200 mA TO
- b) Pulse width: 0 TO +3 VOLTS
- c) Rise time: NA
- d) Fall time: NA
- e) PRF: DC TO 1 MHz
- f) Jitter, stability: OK
- g) Prime power: 120 / 240 V
50-60 Hz

100 mA / DIV
RESISTIVE LOAD
 $R_L = 10 \Omega$
OFFSET = 100 mA
MODULATION INPUT
1 V_{PP} (SLOW RISE TIME
1 MHz SQUARE WAVE)